

CALIBRATION PROTOCOL

STABILOCK

4015

Serial No ..: 0889112

----- test passed -----

Software Version: 1.60 9B74

The instruments used for these measurements are calibrated on a schedule.
The calibration is traceable to national and international standards.

Environmental Conditions:
Relative humidity 30% to 65%
Ambient temperature 23 +-4°C

Warm - up time:
Unit Under Test :1.5h
Test equipment :1.5h

page	Titel	Testsoftware Version	Test equipment
1-1....1-3	RF-Output level/Counter..	3.08P	Stabilock 4032
2-1....2-2	RX-Modulation.....	3.08P	Reference
3-1....3-2	TX-Modulation reading...	3.08P	Serial Number
3-3	Option Duplex	3.08P	
4-1....	RF Offset Counter.....	3.08P	1488263
5-1....5-3	AF-Measurement.....	3.08P	
6-1....6-2	RF-Level reading.....	3.08P	

Authorized tester

Name : Robert KOSCIESZA

Signature:.....

date : 26 Oct 2008